



Nurse Assist LLC Announces a Product Recall

HALTOM CITY, TX, USA, November 6, 2023 /EINPresswire.com/ -- Nurse Assist, LLC Issues Recall of 0.9% Sodium Chloride Irrigation USP and Sterile Water for Irrigation USP Nationwide and to Canada

4409 Haltom Road, Haltom City, TX 71137
800-649-6800/www.nurseassist.com

On November 6, 2023, Nurse Assist, LLC initiated a voluntary recall of 0.9% sodium chloride irrigation USP and sterile water for irrigation USP, due to the potential for a lack of sterility assurance, which could result in a nonsterile product. In populations most at risk, such as patients who are immunocompromised, there is a possibility that the use of the affected product could potentially result in severe or life-threatening adverse events.

To date, Nurse Assist, LLC has not received any reports of adverse events for these products.

The product was distributed in the United States and Canada. Recalled Product was distributed from November 1, 2021 to September 18, 2023.

Recalled products include the following products, these products may be packaged inside kits:

Product	Size
0.9% Sodium Chloride Irrigation USP	100 mL bottles, 250 mL bottles, 500 mL bottles, 1000mL bottles, 3.1oz spray can, 7.1oz spray can, 3mL syringes, 5mL syringes, and 10mL syringes

Sterile Water for Irrigation USP	100 mL bottles, 250 mL bottles, 500 mL bottles, 1000mL bottles, 120 mL cups, 10mL syringes, and 30mL syringes
----------------------------------	---------------------------------------------------------------------------------------------------------------

Expiration dates range from November 1, 2023, to September 18, 2025, for all part numbers, with the exception of part number 1030A (USP STERILE WATER SYRINGE, UDI+B15010304N), which has an expiration date range from November 1, 2023, to September 18, 2028

Detailed product information is as follows:

Brand Name	Part Number	Description
UDI		
NURSE ASSIST	6492775	100ML STERILE SALINE BOTTLE

B15064927754G NURSE ASSIST	6492776	100ML STERILE WATER BOTTLE
B15064927764H NURSE ASSIST	1010	10CC PREFILLED SYRINGES
+B15010104L NURSE ASSIST	51638	10CC PREFILLED SYRINGES
+B150516384% NURSE ASSIST	1210-BP	10ML IV FLUSH SYRINGE
+B1501210BP4G NURSE ASSIST	6496723	30CC PREFILLED SYRINGE
B15064967234D NURSE ASSIST	52333	30CC PREFILLED SYRINGES
+B150523334Z NURSE ASSIST	1203-BP	3ML IV FLUSH SYRINGE
+B1501203BP4I NURSE ASSIST	1205-BP	5ML IV FLUSH SYRINGE
+B150120BP4K NURSE ASSIST	6496722	SYRINGE PREFILLED 10 ML
B15064967224C NURSE ASSIST	1030	USP STERILE WATER SYRINGE
+B15010304N NURSE ASSIST	1030A	USP STERILE WATER SYRINGE
+B1501030A4X STERICARE	6281	1000ML STERILE SALINE BOTTLE
(01)10850421008843 STERICARE	6291	1000ML STERILE WATER BOTTLE
(01)10850421008799 STERICARE	6240	100ML STERILE SALINE BOTTLE
(01)10850421008812 STERICARE	6250	100ML STERILE WATER BOTTLE
(01)10850421008768 STERICARE	6220	120ML STERILE SALINE FOIL LID CUP
(01)10850421008874 STERICARE	6210	120ML STERILE WATER FIOLE LID CUP
(01)10850421008867 STERICARE	6507	210ML SALINE WOUND FLUSH
(01)10850421008744 STERICARE	6270	250ML STERILE SALINE BOTTLE
(01)10850421008829 STERICARE	6260	250ML STERILE WATER BOTTLE
(01)10850421008775 STERICARE	5280	500ML STERILE SALINE BOTTLE
850421008907		

STERICARE (01)10850421008836	6280	500ML STERILE SALINE BOTTLE
STERICARE 850421008914	5290	500ML STERILE WATER BOTTLE
STERICARE (01)10850421008782	6290	500ML STERILE WATER BOTTLE
STERICARE (01)10850421008751	6503	90ML USP SALINE WOULD FLUSH
CARDINAL (01)50192253044490	1020	100ML STERILE SALINE BOTTLE
CARDINAL (01)50192253044506	1022	100ML STERILE SALINE BOTTLE
CARDINAL (01)50192253044506	PT00103427	100ML STERILE SALINE BOTTLE
CARDINAL (01)50192253044513	1024	100ML STERILE WATER BOTTLE
CARDINAL (01)50192253044520	10000	100ML STERILE WATER BOTTLE
CARDINAL (01)50192253044513	PT00103419	100ML STERILE WATER BOTTLE
COVIDIEN (01)10884521139756	210SAL	STERILE WOUND SOLUTION
COVIDIEN (01)20884521140940	90SAL	STERILE WOUND SOLUTION
HALYARD OWENS MINOR (01)20885632389778	SS003	0.9% USP SODIUM CHLORIDE INJECTION
HALYARD OWENS MINOR (01)20885632389761	SS005	0.9% USP SODIUM CHLORIDE INJECTION
HALYARD OWENS MINOR (01)20885632389754	SS010	0.9% USP SODIUM CHLORIDE INJECTION
HALYARD OWENS MINOR (01)20885632395939	49154	1000ML STERILE SALINE BOTTLE
HALYARD OWENS MINOR (01)20885632395892	49150	1000ML STERILE WATER BOTTLE
HALYARD OWENS MINOR (01)20885632395946	49155	100ML STERILE SALINE BOTTLE
HALYARD OWENS MINOR (01)20885632395908	49151	100ML STERILE WATER BOTTLE
HALYARD OWENS MINOR (01)20885632389846	SS010SFR	10ML SYRINGE STERILE
HALYARD OWENS MINOR (01)20885632395915	49152	250ML STERILE SALINE BOTTLE
HALYARD OWENS MINOR	49148	250ML STERILE WATER BOTTLE

(01)20885632395878			
HALYARD OWENS MINOR v	49153	500ML STERILE SALINE BOTTLE	
(01)20885632395922			
HALYARD OWENS MINOR	49149	500ML STERILE WATER BOTTLE	
(01)20885632395885			
IDEXX	98-09444-01	BOTTLE STERILE WATER 10 PACK	No
UDI			
MAC MEDICAL	9210	10ML FILL SYRINGE LABEL	
20850020865042			
MCKESSON	37-6281	1000ML STERILE SALINE BOTTLE	
612479228399			
MCKESSON	37-6291	1000ML STERILE WATER BOTTLE	
612479228405			
MCKESSON	37-6240	100ML STERILE SALINE BOTTLE	
612479168565			
MCKESSON	37-6250	100ML STERILE WATER BOTTLE	
612479168572			
MCKESSON	37-6220	120ML STERILE SALINE CUP	
612479168558			
MCKESSON	37-6210	120ML STERILE WATER CUP	
612479168541			
MCKESSON	37-6507	210ML SALINE WOUND FLUSH	
612479133761			
MCKESSON	37-6270	250ML STERILE SALINE BOTTLE	
612479168596			
MCKESSON	37-6260	250ML STERILE WATER BOTTLE	
612479168589			
MCKESSON	37-6280	500ML STERILE SALINE BOTTLE	
612479180307			
MCKESSON	37-6290	500ML STERILE WATER BOTTLE	
612479180314			
MCKESSON	37-6503	88.7ML SALINE WOUND FLUSH STERILE	
612479228429			
MEDICHOICE OWENS MINOR	SWF071	CAN STERILE SALINE SPRY MEDI 7.1 OZ	
(01)50885632127043			
MEDLINE	DYND1000MLS	1000ML STERILE SALINE BOTTLE	
(01)208882776699466			
MEDLINE	52011	100ML STERILE SALINE BOTTLE	
Customer Kitted			
MEDLINE	RDI30296	100ML STERILE SALINE BOTTLE	
(01)40884389338248			
MEDLINE	54380	100ML STERILE WATER BOTTLE	
Customer Kitted			

MEDLINE (01)40884389338231	RDI30295	100ML STERILE WATER BOTTLE	
MEDLINE (01)20363807109166	EMZ10091240	10ML IV FLUSH STERILE FIELD	
MEDLINE (01)40884389334233	DYND40540	110ML STERILE SALINE CUP	
MEDLINE (01)40884389334226	DYND40570	110ML STERILE WATER CUP	
MEDLINE (01)40884389338262	PCS1650	250ML STERILE SALINE BOTTLE	
MEDLINE (01)40080196427056	PCS1550	250ML STERILE WATER BOTTLE	
MEDLINE (01)40884389118406	DYND500MLS	500ML STERILE SALINE BOTTLE	
MEDLINE (01)40884389118390	DYND500MLW	500ML STERILE WATER BOTTLE	
MEDLINE 40080196297970	CURSALINE7	CAN SALINE STERILE SPRAY CURAD	
SOL (01)20810062438895	PFF001	0.9% SODIUM CHLORIDE IV FLUSH SYRINGE	
TRUDELL UDI	T167007	CUP STERILE SALINE FOIL LID 120 ML	No
TRUDELL UDI	T167005	CUP STERILE WATER FOIL LID 120 ML	No
VYAIRE Discontinued	4873C	CUP STERILE SALINE FOIL LID 120 ML	

Recalled products may be repackaged into medical device convenience kits.

The potential for a compromised sterile barrier was identified during routine product testing. Further investigations prompted a widening of product scope out of an abundance of caution.

Consumers who have these products should stop using them and contact their distributor or Nurse Assist, LLC for return authorization.

Nurse Assist, LLC is notifying its distributors and customers by electronic communication and letter and is arranging for replacement of all recalled products. Users are being requested to coordinate the return of product with Nurse Assist, LLC.

Consumers with questions may contact the company via telephone at 800-649-6800 Monday through Friday between the hours of 8 a.m. and 430 p.m. (CST). Consumers may also contact the company via e-mail at productremovalinfo@nurseassist.com.

Nurse Assist, LLC has notified the FDA of this action.

Adverse reactions or quality problems experienced with the use of this product may be reported to the FDA's MedWatch Adverse Event Reporting program either online, by regular mail, or by fax.

- Complete and submit the report Online: www.fda.gov/medwatch/report.htm
- Regular Mail or Fax: Download the form www.fda.gov/MedWatch/getforms.htm or call 1-800-332-1088 to request a reporting form, then complete and return to the address on the pre-addressed form, or submit by fax to 1-800-FDA-0178

Tom Star
Nurse Assist LLC
tstar@nurseassist.com

This press release can be viewed online at: <https://www.einpresswire.com/article/666610677>

EIN Presswire's priority is source transparency. We do not allow opaque clients, and our editors try to be careful about weeding out false and misleading content. As a user, if you see something we have missed, please do bring it to our attention. Your help is welcome. EIN Presswire, Everyone's Internet News Presswire™, tries to define some of the boundaries that are reasonable in today's world. Please see our Editorial Guidelines for more information.

© 1995-2023 Newsmatics Inc. All Right Reserved.